
INTERFACE / SOFTWARE ET R-3000W/BTA WIN-7 (2.08.3000)



MH1

DURATION: 0:00:20


MINIMUM: 4.75 cN


MAXIMUM: 43.94 cN

STANDARD-DEV.: 5.04 cN

AVERAGE VALUE: 16.08 cN


VARIATION COEFF.:31.33 %





INSTRUMENTS ZÜRICH

TEST:	00	LOT:	ALFA	DATE:	03:29:2000 / 11:26
YARN:	GLASFIBER	COND.:	N	OPTION:	
FILE:	VETROTE1	MISANC.:		OPERATO	OR: a.hall




Graph with "eventmarker" to relate tensions to machine movements

INSTRUMENTS ZÜRICH

MEASURING CYCLE Print-out

- A "dash dot " filter line
- B Zoom section (see following page 3 -)
- C Statistics (Only 40 seconds out of a total of 65 sec.)
- D Histogram No values under the %-setting of the filter line (5 cN)



TIMING:

In addition to the date, the starting time of the Measurements is shown.

INSTRUMENTS ZÜRICH

ZOOM SECTION 10 sec. Print-out (see preceding page 2 section B)

Only the measured data above the Filter-line are considered (6 sec.)

